



This project is funded by
the European Union



Romania-Republic of Moldova
ENI-CROSS BORDER COOPERATION

Subject: Acquisition of special intervention vehicles for specific SMURD missions

Publication Ref: 2020/S 045-107594

RESPONSE TO THE LETTER OF CLARIFICATION

Clarification request	Response to the request
<p>Referring to the Lot. no. 2 “8 units Special vehicles for emergency medical care, first aid, extrication and rescue, type FRAP”, document “TECHNICAL SPECIFICATION - special vehicle for emergency medical care, first aid, extrication and rescue”, point #5.6.20.4 Defibrillator with monitoring functions j) display minimum 5,5 inch diagonal; m) maximum sizes : 245 x 140 x 130 mm (no cover/case or cables included); n) maximum weight : 3 kgs; please also accept defibrillators with 8,5 inch display, size 300 (H) × 310 (W) × 200 (D) mm and 6 kgs weight, in order to allow multiple participants at this tender</p>	<p>The Contracting Authority will not accept a defibrillator with characteristics that do not fall within the limits of the technical specification, pos.5.6.20.4 in Lot no.2 (FRAP special vehicle).</p>

Response letter: nr.19/10-902 from 05.06.2020

Ex.: Dorin MATVEEV
e-mail: logistica@dse.md /
alternative e-mail: dorin.matveev@yahoo.com